



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Fox Valley Metrology, Ltd.
30447 Stacy Pond Drive
Stacy MN 55079

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2005

and national standard

ANSI/NCSL Z540-1-1994 (R2002) &
ANSI/NCSL Z540.3-2006 (R2013)

while demonstrating technical competence in the fields of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations and/or tests to which this accreditation applies.

ACT-1272.03

Certificate Number


ANAB Approval

Certificate Valid: 05/08/2018-06/15/2019

Version No. 002 Issued: 05/08/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005,
ANSI/NCSL Z540-1-1994 (R2002) AND ANSI/NCSL Z540.3-2006 (R2013)**

Fox Valley Metrology, Ltd.
30447 Stacy Pond Drive
Stacy, MN 55079
Chris Kuczynski 920-527-1810

CALIBRATION

Valid to: **June 15, 2019**

Certificate Number: **ACT-1272.03**

Acoustics and Vibration

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Sound Level – Source ¹ 100 Hz, 250 Hz, 500 Hz, 1 000 Hz, 2 000 Hz	114 dB	0.6 dB	Gen Rad 1562-A Sound Level Calibrator

Chemical Quantities

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
pH Meters ¹	(4.01, 7, 10) pH	0.02 pH	pH Buffer Solutions
Conductivity Meters ¹	12.85 mS/cm 1408 µS/cm	0.18 mS/cm 14 µS/cm	Conductivity Solutions
Refractometers ¹	(0.0, 18.0, 29.7) Brix	0.24 Brix	Refractive Index Solutions

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
DC Voltage - Source ¹ Fixed Value	10 V	0.8 µV/V	Fluke 732B Voltage Standard



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
DC Voltage - Source ¹	Up to 220 mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V 220 V to 1.1 kV	12 $\mu\text{V/V} + 0.4 \mu\text{V}$ 5.8 $\mu\text{V/V} + 0.7 \mu\text{V}$ 4.2 $\mu\text{V/V} + 2.5 \mu\text{V}$ 4.1 $\mu\text{V/V} + 4 \mu\text{V}$ 5.8 $\mu\text{V/V} + 40 \mu\text{V}$ 7.6 $\mu\text{V/V} + 0.4 \text{mV}$	Fluke 5720A Multiproduct Calibrator
DC Voltage - Measure ¹	Up to 100 mV 100 mV to 1 V (1 to 10) V (10 to 100) V 100 V to 1 kV	7.8 $\mu\text{V/V} + 0.8 \mu\text{V}$ 5.7 $\mu\text{V/V} + 0.8 \mu\text{V}$ 5.6 $\mu\text{V/V} + 1 \mu\text{V}$ 7.9 $\mu\text{V/V} + 80 \mu\text{V}$ 7.9 $\mu\text{V/V} + 0.15 \text{mV}$	Agilent 3458A Opt 002 Multimeter
DC Voltage - Measure ¹	Up to 200 mV 200 mV to 2 V (2 to 20) V (20 to 200) V 200 V to 1.05 kV	5 $\mu\text{V/V} + 0.10 \mu\text{V}$ 3.5 $\mu\text{V/V} + 0.4 \mu\text{V}$ 3.5 $\mu\text{V/V} + 4 \mu\text{V}$ 5.5 $\mu\text{V/V} + 40 \mu\text{V}$ 5.5 $\mu\text{V/V} + 500 \mu\text{V}$	Fluke 8508A Multimeter
DC High Voltage - Measure ¹	(1 to 10) kV (10 to 100) kV	60 V 0.6 kV	Hipotronics KVM-100 High Voltage Meter
DC Current - Source ¹	Up to 220 μA 220 μA to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mV to 2.2 A	0.12 mA/A + 6 nA 42 $\mu\text{A/A} + 7 \text{nA}$ 41 $\mu\text{A/A} + 40 \text{nA}$ 52 $\mu\text{A/A} + 0.7 \mu\text{A}$ 93 $\mu\text{A/A} + 12 \mu\text{A}$	Fluke 5720A Multiproduct Calibrator
	(2.2 to 11) A (11 to 20.5) A	0.58 mA/A + 0.5 mA 1.2 mA/A + 0.75 mA	Fluke 5520A Multiproduct Calibrator
	(20.5 to 1 000) A	86 mA/A + 0.5 A	Fluke 5520A Multiproduct Calibrator with 50-turn Coil
DC Current - Measure ¹	Up to 100 nA 100 nA to 1 μA (1 to 10) μA (10 to 100) μA 100 μA to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	48 $\mu\text{A/A} + 65 \text{pA}$ 35 $\mu\text{A/A} + 65 \text{pA}$ 35 $\mu\text{A/A} + 0.15 \text{nA}$ 35 $\mu\text{A/A} + 1.3 \text{nA}$ 35 $\mu\text{A/A} + 10 \text{nA}$ 36 $\mu\text{A/A} + 0.1 \mu\text{A}$ 15 $\mu\text{A/A} + 1 \mu\text{A}$ 0.14 mA/A + 20 μA	Agilent 3458A Opt 002 Multimeter



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
DC Current - Measure ¹	Up to 200 μ A 200 μ A to 2 mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A	12 μ A/A + 0.4 nA 12 μ A/A + 4 nA 14 μ A/A + 40 nA 48 μ A/A + 0.8 μ A 0.19 mA/A + 16 μ A 4 mA/A + 0.4 mA	Fluke 8508A Multimeter
DC Current - Measure ¹	(1 to 10) A	2.4 mA/A + 0.7 mA	Fluke DMM
Resistance - Measure ¹ Normal Mode	Up to 2 Ω (2 to 20) Ω (20 to 200) Ω 200 Ω to 2 k Ω (2 to 20) k Ω (20 to 200) k Ω 200 k Ω to 2 M Ω (2 to 20) M Ω (20 to 200) M Ω	17 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ 9.5 $\mu\Omega/\Omega$ + 14 $\mu\Omega$ 8 $\mu\Omega/\Omega$ + 50 $\mu\Omega$ 8 $\mu\Omega/\Omega$ + 0.5 m Ω 8 $\mu\Omega/\Omega$ + 5 m Ω 8 $\mu\Omega/\Omega$ + 50 m Ω 9 $\mu\Omega/\Omega$ + 1 Ω 20 $\mu\Omega/\Omega$ + 0.1 k Ω 0.12 m Ω/Ω + 10 k Ω	Fluke 8508A Multimeter
Resistance - Measure ¹ High Voltage Mode	(2 to 20) M Ω (20 to 200) M Ω 200 m Ω to 2 G Ω (2 to 20) G Ω	17 $\mu\Omega/\Omega$ + 10 Ω 65 $\mu\Omega/\Omega$ + 1 k Ω 0.18 m Ω/Ω + 0.1 M Ω 15 m Ω/Ω + 10 M Ω	Fluke 8508A Multimeter
Resistance - Source ¹	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω	0.11 m Ω 0.11 m Ω 0.21 m Ω 0.27 m Ω 0.51 m Ω 1.4 m Ω 2.6 m Ω 11 m Ω 21 m Ω 0.11 Ω 0.21 Ω 1.3 Ω 2.7 Ω 24 Ω 48 Ω 0.48 k Ω	Fluke 5720A Multiproduct Calibrator



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Resistance - Source ¹	19 MΩ 100 MΩ	1.1 kΩ 23 kΩ	Fluke 5720A Multiproduct Calibrator
	1 GΩ 10 GΩ 100 GΩ	1.9 MΩ 47 MΩ 0.95 GΩ	IET Labs HRRS Decade Box
Resistance - Measure ¹	Up to 10 Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1 GΩ	24 μΩ/Ω + 0.1 mΩ 20 μΩ/Ω + 1 mΩ 18 μΩ/Ω + 1 mΩ 18 μΩ/Ω + 10 mΩ 18 μΩ/Ω + 0.1 Ω 24 μΩ/Ω + 7 mΩ 87 μΩ/Ω + 0.2 Ω 0.73 mΩ/Ω + 2 Ω 7.2 mΩ/Ω + 20 kΩ	Agilent 3458A Opt 002 Multimeter
AC Voltage - Source ¹	Up to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	2.4 mV/V + 4 μV 2.4 mV/V + 4 μV 2.2 mV/V + 4 μV 2.2 mV/V + 4 μV 2.3 mV/V + 5 μV 2.5 mV/V + 10 μV 2.7 mV/V + 20 μV 3.8 mV/V + 20 μV 0.61 mV/V + 4 μV 0.56 mV/V + 4 μV 0.36 mV/V + 4 μV 0.42 mV/V + 4 μV 0.70 mV/V + 5 μV 1.3 mV/V + 10 μV 1.7 mV/V + 20 μV 3.4 mV/V + 20 μV	Fluke 5720A Multiproduct Calibrator



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage - Source ¹	(22 to 220) mV	0.29 mV/V + 12 μV	Fluke 5720A Multiproduct Calibrator
	(10 to 20) Hz	0.13 mV/V + 7 μV	
	(20 to 40) Hz	0.11 mV/V + 7 μV	
	40 Hz to 20 kHz	0.24 mV/V + 7 μV	
	(20 to 50) kHz	0.54 mV/V + 17 μV	
	(50 to 100) kHz	1.1 mV/V + 20 μV	
	(100 to 300) kHz	1.6 mV/V + 25 μV	
	(300 to 500) kHz	3.3 mV/V + 45 μV	
	500 kHz to 1 MHz		
	(0.22 to 2.2) V	0.28 mV/V + 40 μV	
	(10 to 20) Hz	0.11 mV/V + 15 μV	
	(20 to 40) Hz	55 μV/V + 8 μV	
	40 Hz to 20 kHz	0.12 mV/V + 10 μV	
	(20 to 50) kHz	0.13 mV/V + 30 μV	
	(50 to 100) kHz	0.49 mV/V + 80 μV	
	(100 to 300) kHz	1.2 mV/V + 0.2 mV	
	(300 to 500) kHz	2.0 mV/V + 0.3 mV	
	500 kHz to 1 MHz		
	(2.2 to 22) V	0.28 mV/V + 0.2 mV	
	(10 to 20) Hz	0.11 mV/V + 0.15 mV	
	(20 to 40) Hz	56 μV/V + 50 μV	
	40 Hz to 20 kHz	0.12 mV/V + 0.1 mV	
	(20 to 50) kHz	0.12 mV/V + 0.2 mV	
	(50 to 100) kHz	0.32 mV/V + 0.6 mV	
	(100 to 300) kHz	1.2 mV/V + 2 mV	
	(300 to 500) kHz	1.8 mV/V + 3.2 mV	
	500 kHz to 1 MHz		
	(22 to 220) V	0.28 mV/V + 4 mV	
	(10 to 20) Hz	0.11 mV/V + 1.5 mV	
	(20 to 40) Hz	65 μV/V + 0.6 mV	
40 Hz to 20 kHz	0.12 mV/V + 1 mV		
(20 to 50) kHz	0.18 mV/V + 2.5 mV		
(50 to 100) kHz	1.1 mV/V + 16 mV		
(100 to 300) kHz	5.1 mV/V + 40 mV		
(300 to 500) kHz	9.3 mV/V + 80 mV		
500 kHz to 1 MHz			
220 V to 1.1 kV	0.35 mV/V + 16 mV		
(15 to 50) Hz	88 μV/V + 3.5 mV		
50 Hz to 1 kHz			



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage Harmonics – Source (2 nd to 50 th) ¹ (10 to 45) Hz (45 to 65) Hz (65 to 500) Hz 500 Hz to 5 kHz (5 to 10) kHz	32 mV to 33 V 33 mV to 1 kV 33 mV to 1 kV 330 mV to 1 kV 3.3 V to 1 kV	0.35 mV/V + 16 μV 0.21 mV/V + 16 μV 0.21 mV/V + 16 μV 0.21 mV/V + 0.21 mV 0.21 mV/V + 1.2 mV	Fluke 5520A Multiproduct Calibrator
	Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz 100 mV to 1 V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	0.46 mV/V + 13 μV 0.35 mV/V + 11 μV 0.46 mV/V + 11 μV 1.3 mV/V + 11 μV 5.9 mV/V + 11 μV 46 μV/V + 12 μV 0.14 mV/V + 4.5 μV 0.14 mV/V + 2.5 μV 0.22 mV/V + 2.5 μV 0.41 mV/V + 2.5 μV 0.99 mV/V + 2.5 μV 3.5 mV/V + 11 μV 12 mV/V + 11 μV 18 mV/V + 11 μV 0.14 mV/V + 45 μV 0.14 mV/V + 25 μV 0.22 mV/V + 25 μV 0.41 mV/V + 25 μV 0.99 mV/V + 25 μV 3.5 mV/V + 0.11 mV 12 mV/V + 0.11 mV 18 mV/V + 0.11 mV	



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage - Measure ¹ Bandwidth < 2 MHz	(1 to 10) V		Agilent 3458A Opt 002 Multimeter
	(1 to 40) Hz	0.14 mV/V + 0.45 mV	
	40 Hz to 1 kHz	0.14 mV/V + 0.25 mV	
	(1 to 20) kHz	0.22 mV/V + 0.25 mV	
	(20 to 50) kHz	0.41 mV/V + 0.25 mV	
	(50 to 100) kHz	0.98 mV/V + 0.25 mV	
	(100 to 300) kHz	3.5 mV/V + 1.1 mV	
	300 kHz to 1 MHz	12 mV/V + 1.1 mV	
	(1 to 2) MHz	18 mV/V + 1.1 mV	
	(10 to 100) V		
	(1 to 40) Hz	0.29 mV/V + 4.5 mV	
	40 Hz to 1 kHz	0.29 mV/V + 2.5 mV	
	(1 to 20) kHz	0.29 mV/V + 2.5 mV	
	(20 to 50) kHz	0.29 mV/V + 2.5 mV	
(50 to 100) kHz	1.5 mV/V + 2.5 mV		
(100 to 300) kHz	4.7 mV/V + 11 mV		
300 kHz to 1 MHz	18 mV/V + 11 mV		
100 V to 1 kV	(1 to 40) Hz	0.52 mV/V + 45 mV	
	40 Hz to 1 kHz	0.52 mV/V + 25 mV	
	(1 to 20) kHz	0.75 mV/V + 25 mV	
	(20 to 50) kHz	1.5 mV/V + 25 mV	
	(50 to 100) kHz	3.5 mV/V + 25 mV	
(1 to 10) kV	(50 to 60) Hz	0.12 kV	Hipotronics KVM-100 High Voltage Meter
	(10 to 100) kV		
AC Voltage – Measure ¹ Bandwidth > 2 MHz	(50 to 60) Hz	1.2 kV	
	Up to 10 mV		Agilent 3458A Opt 002 Multimeter
	45 Hz to 100 kHz	1.2 mV/V + 6 μV	
	100 kHz to 1 MHz	14 mV/V + 5.1 μV	
	(1 to 4) MHz	83 mV/V + 7.1 μV	
	(4 to 8) MHz	0.24 V/V + 8.1 μV	
	(10 to 100) mV		
	45 Hz to 100 kHz	1.1 mV/V + 61 μV	
	100 kHz to 1 MHz	24 mV/V + 51 μV	
	(1 to 4) MHz	47 mV/V + 71 μV	
	(4 to 8) MHz	47 mV/V + 81 μV	
	(8 to 10) MHz	0.18 V/V + 0.1 mV	



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage – Measure ¹ Bandwidth > 2 MHz	100 mV to 1 V 45 Hz to 100 kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (8 to 10) MHz	1.1 mV/V + 0.61 mV 24 mV/V + 0.51 mV 47 mV/V + 0.71 mV 47 mV/V + 0.81 mV 0.18 V/V + 1 mV	Agilent 3458A Opt 002 Multimeter
	(1 to 10) V 45 Hz to 100 kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (8 to 10) MHz	1.2 mV/V + 6.1 μV 24 mV/V + 5.1 μV 47 mV/V + 7.1 μV 47 mV/V + 8.1 μV 0.18 V/V + 10 μV	
	(10 to 100) V 45 Hz to 100 kHz	1.5 mV/V + 2.5 mV	Hipotronics KVM-100 High Voltage Meter
	100 V to 1 kV 45 Hz to 100 kHz	3.6 mV/V + 0.11 V	
AC Voltage - Measure ¹ Bandwidth < 1 MHz	Up to 200 mV (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz 200 mV to 2 V (1 to 10) Hz (10 to 40) Hz (40 to 100) Hz 100 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.17 mV/V + 14 μV 0.14 mV/V + 4 μV 0.12 mV/V + 4 μV 0.11 mV/V + 2 μV 0.14 mV/V + 4 μV 0.34 mV/V + 8 μV 0.77 mV/V + 20 μV 0.15 mV/V + 0.12 mV 0.12 mV/V + 20 μV 90 μV/V + 20 μV 75 μV/V + 20 μV 0.11 mV/V + 20 μV 0.22 mV/V + 0.84 mV 0.57 mV/V + 0.2 mV 3 mV/V + 2 mV 10 mV/V + 2 mV	Fluke 8508A Multimeter



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Voltage - Measure ¹ Bandwidth < 1 MHz	(2 to 20) V		Fluke 8508A Multimeter
	(1 to 10) Hz	0.15 mV/V + 1.2 mV	
	(10 to 40) Hz	0.12 mV/V + 0.2 mV	
	(40 to 100) Hz	90 μV/V + 0.2 mV	
	100 Hz to 2 kHz	75 μV/V + 0.2 mV	
	(2 to 10) kHz	0.11 mV/V + 0.2 mV	
	(10 to 30) kHz	0.22 mV/V + 8.4 mV	
	(30 to 100) kHz	0.57 mV/V + 2 mV	
	(100 to 300) kHz	3 mV/V + 20 mV	
	300 kHz to 1 MHz	10 mV/V + 20 mV	
	(20 to 200) V		
	(1 to 10) Hz	0.15 mV/V + 12 mV	
	(10 to 40) Hz	0.12 mV/V + 2 mV	
	(40 to 100) Hz	90 μV/V + 2 mV	
	100 Hz to 2 kHz	75 μV/V + 2 mV	
	(2 to 10) kHz	0.11 mV/V + 2 mV	
	(10 to 30) kHz	0.22 mV/V + 84 mV	
	(30 to 100) kHz	0.57 mV/V + 20 mV	
(100 to 300) kHz	3 mV/V + 0.2 V		
300 kHz to 1 MHz	10 mV/V + 0.2 V		
AC Current - Source ¹	200 V to 1.05 kV		Fluke 5720A Multiproduct Calibrator
	(1 to 10) Hz	0.15 mV/V + 70 mV	
	(10 to 40) Hz	0.12 mV/V + 20 mV	
	40 Hz to 10 kHz	0.12 mV/V + 20 mV	
	(10 to 30) kHz	0.23 mV/V + 40 mV	
	(30 to 100) kHz	0.58 mV/V + 0.2 V	
	Up to 220 μA		
	(10 to 20) Hz	0.30 mA/A + 16 nA	
	(20 to 40) Hz	0.20 mA/A + 10 nA	
	40 Hz to 1 kHz	0.16 mA/A + 8 nA	
(1 to 5) kHz	0.22 mA/A + 12 nA		
(5 to 10) kHz	1.3 mA/A + 65 nA		
220 μA to 2.2 mA			
(10 to 20) Hz	0.31 mA/A + 40 nA		
(20 to 40) Hz	0.22 mA/A + 35 nA		
40 Hz to 1 kHz	0.15 mA/A + 35 nA		
(1 to 5) kHz	0.24 mA/A + 0.11 μA		
(5 to 10) kHz	1.3 mA/A + 0.65 μA		



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment	
AC Current - Source ¹	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 mA to 2.2 A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2 to 3) A (10 to 45) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.32 mA/A + 0.4 μA 0.23 mA/A + 0.35 μA 0.15 mA/A + 0.35 μA 0.24 mA/A + 0.55 μA 1.3 mA/A + 5 μA 0.30 mA/A + 4 μA 0.20 mA/A + 3.5 μA 0.15 mA/A + 2.5 μA 0.24 mA/A + 3.5 μA 1.3 mA/A + 10 μA 0.31 mA/A + 35 μA 0.53 mA/A + 80 μA 8.1 mA/A + 0.16 mA 2.1 mA/A + 0.1 mA 0.75 mA/A + 0.1 mA 6.9 mA/A + 1 mA 29 mA/A + 5 mA	Fluke 5720A Multiproduct Calibrator	
AC Current - Source ¹	(3 to 11) A (45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz (11 to 20.5) A (45 to 100) Hz (0.1 to 1) kHz (1 to 5) kHz	0.74 mA/A + 2 mA 1.2 mA/A + 2 mA 35 mA/A + 2 mA 1.4 mA/A + 5 mA 1.8 mA/A + 5 mA 35 mA/A + 5 mA	Fluke 5520A Multiproduct Calibrator	
AC Current - Source ¹	(20.5 to 1 000) A (45 to 65) Hz (20.5 to 150) A (65 to 440) Hz	90 mA/A + 0.5 A 0.55 mA/A + 0.5 mA	Fluke 5520A Multiproduct Calibrator w/ 50-turn Coil	
AC Current Harmonics - Source ¹ (2 nd to 50 th)	(10 to 45) Hz (45 to 65) Hz (65 to 500) Hz 500 Hz to 5 kHz (5 to 10) kHz	3.3 mA to 3 A 3.3 mA to 20.5 A 33 mA to 20.5 A 33 mA to 20.5 A (33 to 330) mA	1.1 mA/A + 4 μA 0.5 mA/A + 4 μA 1.2 mA/A + 0.1 mA 2.3 mA/A + 0.2 mA 4.6 mA/A + 0.4 mA	Fluke 5520A Multiproduct Calibrator



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Current - Measure ¹	Up to 100 μ A		Agilent 3458A Opt 002 Multimeter
	(10 to 20) Hz	4.8 mA/A + 30 nA	
	(20 to 45) Hz	1.9 mA/A + 30 nA	
	(45 to 100) Hz	0.83 mA/A + 30 nA	
	100 Hz to 5 kHz	0.83 mA/A + 30 nA	
	100 μ A to 1 mA		
	(10 to 20) Hz	4.9 mA/A + 0.2 μ A	
	(20 to 45) Hz	1.9 mA/A + 0.2 μ A	
	(45 to 100) Hz	0.83 mA/A + 0.2 μ A	
	100 Hz to 5 kHz	0.47 mA/A + 0.2 μ A	
	(5 to 20) kHz	0.83 mA/A + 0.2 μ A	
	(20 to 50) kHz	4.9 mA/A + 0.4 μ A	
	(50 to 100) kHz	6.6 mA/A + 1.5 μ A	
	(1 to 10) mA		
	(10 to 20) Hz	4.9 mA/A + 2 μ A	
	(20 to 45) Hz	1.9 mA/A + 2 μ A	
	(45 to 100) Hz	0.83 mA/A + 2 μ A	
	100 Hz to 5 kHz	0.47 mA/A + 2 μ A	
	(5 to 20) kHz	0.83 mA/A + 2 μ A	
	(20 to 50) kHz	4.9 mA/A + 4 μ A	
	(50 to 100) kHz	6.6 mA/A + 15 μ A	
	(10 to 100) mA		
	(10 to 20) Hz	4.9 mA/A + 20 μ A	
	(20 to 45) Hz	1.9 mA/A + 20 μ A	
	(45 to 100) Hz	0.83 mA/A + 20 μ A	
	100 Hz to 5 kHz	0.47 mA/A + 20 μ A	
	(5 to 20) kHz	0.47 mA/A + 20 μ A	
(20 to 50) kHz	4.9 mA/A + 40 μ A		
(50 to 100) kHz	6.6 mA/A + 0.15 mA		
100 mA to 1 A			
(10 to 20) Hz	4.8 mA/A + 0.2 mA		
(20 to 45) Hz	2 mA/A + 0.2 mA		
(45 to 100) Hz	1.1 mA/A + 0.2 mA		
100 Hz to 5 kHz	1.3 mA/A + 0.2 mA		
(5 to 20) kHz	3.7 mA/A + 0.2 mA		
(20 to 50) kHz	12 mA/A + 0.4 mA		
AC Current - Measure ¹	(1 to 10) A		Fluke DMM
	(20 to 50) Hz	0.23 A/A + 10 mA	
	50 Hz to 2 kHz	36 mA/A + 10 mA	

Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
AC Current - Measure ¹	Up to 200 μ A (1 to 10) Hz	0.31 mA/A + 20 nA	Fluke 8508A Multimeter
	10 Hz to 10 kHz (10 to 30) kHz	0.3 mA/A + 20 nA	
	(30 to 100) kHz	0.71 mA/A + 20 nA	
	200 μ A to 2 mA (1 to 10) Hz	4 mA/A + 20 nA	
	10 Hz to 10 kHz (10 to 30) kHz	0.31 mA/A + 0.2 μ A	
	(30 to 100) kHz	0.3 mA/A + 0.2 μ A	
	(2 to 20) mA (1 to 10) Hz	0.71 mA/A + 0.2 μ A	
	10 Hz to 10 kHz (10 to 30) kHz	4 mA/A + 0.2 μ A	
	(30 to 100) kHz	0.31 mA/A + 2 μ A	
	(20 to 200) mA (1 to 10) Hz	0.3 mA/A + 2 μ A	
	10 Hz to 10 kHz (10 to 30) kHz	0.71 mA/A + 2 μ A	
	(30 to 100) kHz	4 mA/A + 2 μ A	
Capacitance - Measure ¹ 42 Hz to 5 MHz	0.32 pF to 370 mF	1.1 mF/F	Hioki 3532-50 LCR Meter
Capacitance - Source ¹	130 pF to 3.3 nF (3.3 to 11) nF	5.8 mF/F + 10 pF	Fluke 5520A Multiproduct Calibrator
	(11 to 110) nF (110 to 330) nF	2.9 mF/F + 10 pF	
	330 nF to 1.1 μ F (1.1 to 3.3) μ F	2.9 mF/F + 0.1 nF	
	(3.3 to 11) μ F (11 to 33) μ F	2.9 mF/F + 0.3 nF	
	(33 to 110) μ F (110 to 330) μ F	2.9 mF/F + 1 nF	
	330 μ F to 1.1 mF (1.1 to 3.3) mF	2.9 mF/F + 3 nF	
		2.9 mF/F + 10 nF	
		4.7 mF/F + 30 nF	
		5.3 mF/F + 0.1 μ F	
		1 mF/F + 0.3 μ F	
		6 mF/F + 1 μ F	
		5.3 mF/F + 3 μ F	



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Capacitance - Source ¹	(3.3 to 11) mF (11 to 33) mF (33 to 110) mF	5.3 mF/F + 10 μF 8.9 mF/F + 30 μF 13 mF/F + 0.1 mF	Fluke 5520A Multiproduct Calibrator
Phase - Measure ¹	(0 to 360) ° 10 Hz to 2 kHz (2 to 5) kHz (5 to 10) kHz (10 to 50) kHz (50 to 60) kHz (60 to 70) kHz (70 to 80) kHz (80 to 90) kHz (90 to 100) kHz (100 to 500) kHz 500 kHz to 1 MHz	0.026 ° 0.036 ° 0.048 ° 0.059 ° 0.07 ° 0.082 ° 0.093 ° 0.1 ° 0.12 ° 0.58 ° 1.2 °	Clark Hess 6000A Phase Meter
Oscilloscopes ¹ DC Voltage (50 Ω) DC Voltage (1 MΩ) AC Voltage (50 Ω) AC Voltage (1 MΩ) Leveled Sine Wave 50 kHz to 1.1 GHz Time Markers Wave Generator (50 Ω) Wave Generator (1 MΩ) Pulse Generator - Width Pulse Generator - Period Input Impedance Measure	1 mV to 6.6 V 1 mV to 130 V 1 mV to 6.6 V 1 mV to 130 V 5 mV to 5.5 V 1 ns to 5 s 1.8 mV to 2.5 V p-p 1.8 mV to 55 V p-p (4 to 45) nS (45 to 500) nS 200 ns to 20 mS (50 to 60) Ω 500 kΩ to 1 MΩ	2.9 mV/V + 40 μV 0.55 mV/V + 40 μV 2.9 mV/V + 40 μV 1.1 mV/V + 40 μV 51 mV/V + 0.1 mV 6.4 μs/s 35 mV/V + 0.10 mV 35 mV/V + 0.10 mV 58 mS/S + 0.5 ns 58 mS/S + 4 ns 58 mS/S + 0.2 μs 1.2 mΩ/Ω 1.2 mΩ/Ω	Fluke 5520A SC1100 Multiproduct Calibrator



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
DC Power - Source ¹	10 mW to 330 W 330 W to 3 kW (3 to 20.5) kW	0.27 mW/W 0.26 mW/W 0.82 mW/W	Fluke 5520A Multiproduct Calibrator
AC Power - Source ¹	100 μW to 9 W (9 to 33) W (33 to 90) W (90 to 330) W (330 to 900) W 900 W to 2.2 kW	1.7 mW/W 1.2 mW/W 1.7 mW/W 1.2 mW/W 11 mW/W 4.6 mW/W	Fluke 5520A Multiproduct Calibrator
Electrical Simulation of RTDs ¹	Pt 385, 100 Ω (-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (500 to 630) °C (630 to 800) °C Pt 3926, 100 Ω (-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (500 to 630) °C Pt 3916 (JIS) 100 Ω (-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.05 °C 0.05 °C 0.07 °C 0.09 °C 0.10 °C 0.12 °C 0.23 °C 0.05 °C 0.05 °C 0.07 °C 0.09 °C 0.10 °C 0.12 °C 0.25 °C 0.04 °C 0.05 °C 0.06 °C 0.07 °C 0.09 °C 0.10 °C 0.23 °C	Fluke 5520A Multiproduct Calibrator



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Electrical Simulation of RTDs ¹	Pt 385, 200 Ω		Fluke 5520A Multiproduct Calibrator
	(-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.12 °C	
	(300 to 400) °C	0.13 °C	
	(400 to 600) °C	0.14 °C	
	(600 to 630) °C	0.16 °C	
	Pt 385, 500 Ω		
	(-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 260) °C	0.06 °C	
	(260 to 300) °C	0.08 °C	
	(300 to 400) °C	0.08 °C	
	(400 to 600) °C	0.09 °C	
	(600 to 630) °C	0.11 °C	
	Pt 385, 1 000 Ω		
	(-200 to -80) °C	0.03 °C	
	(-80 to 0) °C	0.03 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
(260 to 300) °C	0.06 °C		
(300 to 400) °C	0.07 °C		
(400 to 600) °C	0.07 °C		
(600 to 630) °C	0.23 °C		
PtNi 385, 120 Ω, Ni 120			
(-80 to 0) °C	0.08 °C		
(0 to 100) °C	0.08 °C		
(100 to 260) °C	0.14 °C		
Cu 427, 10 Ω			
(-100 to 260) °C	0.03 °C		
Electrical Simulation of Thermocouples ¹	Type K		Fluke 5520A Multiproduct Calibrator
	(-200 to -100) °C	0.33 °C	
	(-100 to -25) °C	0.18 °C	
	(-25 to 120) °C	0.16 °C	
	(120 to 1 000) °C	0.26 °C	
	(1 000 to 1 372) °C	0.40 °C	



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Electrical Simulation of Thermocouples ¹	Type J		Fluke 5520A Multiproduct Calibrator
	(-210 to -100) °C	0.27 °C	
	(-100 to -30) °C	0.16 °C	
	(-30 to 150) °C	0.14 °C	
	(150 to 760) °C	0.17 °C	
	(760 to 1 200) °C	0.23 °C	
	Type E		
	(-250 to -100) °C	0.50 °C	
	(-100 to -35) °C	0.16 °C	
	(-25 to 350) °C	0.14 °C	
	(350 to 650) °C	0.16 °C	
	(650 to 1 000) °C	0.21 °C	
	Type T		
	(-250 to -150) °C	0.63 °C	
	(-150 to 0) °C	0.24 °C	
	(0 to 120) °C	0.16 °C	
	(120 to 400) °C	0.14 °C	
	Type S		
	(0 to 250) °C	0.47 °C	
	(250 to 1 000) °C	0.36 °C	
	(1 000 to 1 400) °C	0.37 °C	
	(1 400 to 1 767) °C	0.46 °C	
	Type B		
	(600 to 800) °C	0.44 °C	
	(-100 to -25) °C	0.34 °C	
	(-25 to 120) °C	0.30 °C	
	(120 to 1 000) °C	0.33 °C	
	Type C		
(0 to 150) °C	0.30 °C		
(150 to 650) °C	0.26 °C		
(650 to 1 000) °C	0.31 °C		
(1 000 to 1 800) °C	0.50 °C		
(1 800 to 2 316) °C	0.84 °C		
Type L			
(-200 to -100) °C	0.37 °C		
(-100 to 800) °C	0.26 °C		
(800 to 900) °C	0.17 °C		



Electrical – DC/Low Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Electrical Simulation of Thermocouples ¹	Type N (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1 300) °C Type R (0 to 250) °C (250 to 400) °C (400 to 1 000) °C (1 000 to 1 767) °C Type U (-200 to 0) °C (0 to 600) °C	0.40 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C 0.57 °C 0.35 °C 0.33 °C 0.40 °C 0.56 °C 0.27 °C	Fluke 5520A Multiproduct Calibrator
Inductance - Source ¹	(1 to 10) mH (10 to 100) mH 100 mH to 1 H (1 to 10) H	22 mH/H 11 mH/H 6 mH/H 3 mH/H	General Radio 1490-D Decade Inductor
Ionizers ¹ Decay Time Float Voltage	(0.1 to 999.9) s (-1 100 to 1 100) V	0.2 s 3.1 V	Trek 156A Charged Plate Monitor

Electrical - RF/Microwave

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
RF Power - Measure ^{1,4} Absolute Level 100 kHz to 3 GHz (3 to 18) GHz (18 to 26.5) GHz	(+20 to +30) dBm	0.37 dB 0.39 dB 0.4 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules
100 kHz to 3 GHz (3 to 18) GHz (18 to 26.5) GHz	(-20 to +20) dBm	0.15 dB 0.18 dB 0.21 dB	



Electrical - RF/Microwave

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Amplitude Modulation - Source ^{1,4} Rate: DC to 100 kHz Depths: 0 % to 100 %	250 kHz to 40 GHz	7.1 % of setting + 1 %	Agilent E8257D Signal Generator
Amplitude Modulation - Measure ^{1,4} 100 kHz to 10 MHz 10 MHz to 3 GHz 10 MHz to 3 GHz (3 to 26.5) GHz (3 to 26.5) GHz	Rate: 50 Hz to 10 kHz Depths: 5 % to 99 % Rate: 50 Hz to 100 kHz Depths: 20 % to 99 % Rate: 50 Hz to 100 kHz Depths: 5 % to 20 % Rate: 50 Hz to 100 kHz Depths: 20 % to 99 % Rate: 50 Hz to 100 kHz Depths: 5 % to 20 %	2.2 % of reading 1.2 % of reading 4.2 % of reading 3.5 % of reading 6 % of reading	Agilent N5531S Measuring Receiver with N5532A Sensor Modules
Phase Modulation - Source ^{1,4} Rate: DC to 100 kHz	250 kHz to 40 GHz	5.9 % setting + 0.1 rad	Agilent E8257D Signal Generator
Phase Modulation - Measure ^{1,4} 100 kHz to 6.6 GHz 100 kHz to 6.6 GHz (6.6 to 13.2) GHz (6.6 to 13.2) GHz (13.2 to 26.5) GHz (13.2 to 26.5) GHz	Rate: 200 Hz 20 kHz Dev.: > 0.7 rad Rate: 200 Hz 20 kHz Dev.: > 0.3 rad Rate: 200 Hz 20 kHz Dev.: > 2.0 rad Rate: 200 Hz 20 kHz Dev.: > 0.6 rad Rate: 200 Hz 20 kHz Dev.: > 2.0 rad Rate: 200 Hz 20 kHz Dev.: > 0.6 rad	1.2 % of reading 3.6 % of reading 1.2 % of reading 3.6 % of reading 1.2 % of reading 3.6 % of reading	Agilent N5531S Measuring Receiver with N5532A Sensor Modules



Electrical - RF/Microwave

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Tuned RF Level - Measure ^{1,4} Absolute Level 500 kHz to 3.05 GHz (3.05 to 6.6) GHz (6.6 to 13.2) GHz (13.2 to 19.2) GHz (19.2 to 26.5) GHz	(+16 to +30) dBm	0.37 dB + 0.005 dB/10 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules
	(-106 to +16) dBm	0.15 dB + 0.005 dB/10 dB	
	(-129 to -106) dBm	0.15 dB + 0.12 dB/10 dB	
	(+20 to +30) dBm	0.39 dB + 0.005 dB/10 dB	
	(-90 to +20) dBm	0.18 dB + 0.005 dB/10 dB	
	(-114 to -90) dBm	0.23 dB + 0.12 dB/10 dB	
	(+20 to +30) dBm	0.39 dB + 0.005 dB/10 dB	
	(-81 to +20) dBm	0.18 dB + 0.005 dB/10 dB	
	(-104 to -81) dBm	0.23 dB + 0.12 dB/10 dB	
	(+20 to +30) dBm	0.4 dB + 0.005 dB/10 dB	
	(-70 to +20) dBm	0.21 dB + 0.005 dB/10 dB	
	(-93 to -70) dBm	0.25 dB + 0.12 dB/10 dB	
	(+20 to +30) dBm	0.4 dB + 0.005 dB/10 dB	
	(-62 to +20) dBm	0.21 dB + 0.005 dB/10 dB	
	(-85 to -62) dBm	0.24 dB + 0.12 dB/10 dB	
Tuned RF Level - Measure ^{1,4} Relative Level 500 kHz to 3.05 GHz	(-90 to +30) dBm	0.026 dB + 0.005 dB/10 dB	Agilent N5531S Measuring Receiver with N5532A Sensor Modules
	(-106 to -90) dBm	0.067 dB + 0.12 dB/10 dB	
	(-129 to -106) dBm	0.076 dB + 0.12 dB/10 dB	
RF Power Sensors- Calibration Factor ^{1,4} 100 kHz to 10 MHz 10 MHz to 10 GHz (10 to 18) GHz	(-20 to +14) dBm	1.5 %	Tegam 1827 Power Sensor Calibrator, Agilent 3458A Multimeter, Agilent E8257D Signal Generator, Agilent E4419B Power Meter, Agilent 3325B Function Generator
		1.5 %	
		1.7 %	



Electrical - RF/Microwave

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Frequency Modulation - Measure ^{1,4} 250 kHz to 10 MHz 10 MHz to 3 GHz	Rate: 20 Hz to 10 kHz Dev.: ≤ 40 kHz peak Rate: 20 Hz to 200 kHz Dev.: ≤ 400 kHz peak	3.1 % of reading 3.1 % of reading	Agilent N5531S Measuring Receiver with N5532A Sensor Modules
Frequency Modulation - Measure ^{1,4} (3 to 26.5) GHz	Rate: 20 Hz to 200 kHz Dev.: ≤ 400 kHz peak	7.7 % of reading	Agilent N5531S Measuring Receiver with N5532A Sensor Modules
Frequency Modulation - Source ^{1,4} 250 kHz to 40 GHz	1 dB Rate: DC to 100 kHz 3 dB Rate: DC to 10 MHz Dev.: ≤ (N X 800 kHz)	4.2 % setting + 20 Hz	Agilent E8257D Signal Generator
Pulse Generation - Measure ^{1,4} DC to 225 MHz Pulse Width Rise/Fall Time	5 ns to 10 ⁵ s 5 ns to 10 ⁵ s	1.1 ns 1.1 ns	Agilent 53132A Counter
Pulse Generation - Source ^{1,4} Repetition Frequency: 0.024 Hz to 14.28 MHz Period: 70 ns to 42 s	10 ns to 42 s	17 ns	Agilent E8257D Signal Generator
Pulse Modulation - Source ^{1,4} Level Rise/Fall Time	(0 to 9) dBm 10 MHz to 40 GHz	0.59 dBm 12 ns	Agilent E8257D Signal Generator

Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Cylindrical Rings ^{1,2}	(0.25 to 8) in	(12 + 3D) μin	Fowler Lab Concept Measuring Machine ASME B89.1.6
Cylindrical Plugs ^{1,2}	(0.010 to 4) in	(53 + 0.4D) μin	Plug gage Comparator



Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Threaded Plugs ^{1,2} Pitch Diameter Major Diameter	(0.01 to 4) in (0.01 to 4) in	(73 + 3.2D) μin (53 + 4.1D) μin	Plug gage Comparator
Calipers ^{1,2}	Up to 80 in	(380 + 15L) μin	Gage Blocks
Indicators ^{1,2}	Up to 4 in	(36 + 10L) μin	Indicator Checker
Test Indicators ¹	Up to 0.06 in	39 μin	Indicator Checker
OD Micrometers ^{1,2}	Up to 60 in	(72 + 12L) μin	Gage Blocks
ID Micrometer ^{1,2}	(1.5 to 40) in	(370 + 7L) μin	Gage Blocks
Height Gages ^{1,2}	Up to 40 in	(96 + 14L) μin	Gage Blocks
Bore Gages ¹	(0.25 to 12) in	350 μin	Cylindrical Rings
Crimpers ¹ Die Check Crimp Height	(0.011 to 0.5) in (0.01 to 0.5) in	230 μin 0.001 2 in	Pin Gages Micrometer
Profilometers ¹ Ra	(2 to 300) μin	2.2 μin	Roughness Specimen
Surface Plates ^{1,2} Repeat Reading Overall Flatness	(4 to 34) in Diagonal (34 to 175) in Diagonal	(30 + 0.2D) μin (66 + 0.2D) μin	Repeat – O – Meter Electronic Levels
CMM Calibration ^{1,2} Volumetric Linearity Linearity	(5 to 40) in (1 to 60) in Above 60 in	(12 + 14L) μin (7 + 14L) μin (20 + 0.4L) μin	Ball Bars Step Gage Renishaw Laser System B89.4.1
Linear Measurements ^{1,2}	Up to 1 560 in	(38 + 0.5L) μin	Laser
Optical Comparators ^{1,2} Linearity Magnification	Up to 12 in 10x, 20x, 31.25x, 50x, 62.5x, 100x, 200x	(97 + 12L) μin 0.000 46 in	Glass Scale Precision Balls Calibration Sphere
Roundness Testers ¹ Axial Error Radial Error	(-1 000 to 1 000) μm	0.15 μm 0.15 μm	Test Sphere
ULMs ¹ Length	(1 to 100) mm	0.19 μm	Gage Blocks



Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Film Thickness Gages ¹	(0.01 to 0.06) in	380 μin	Film Thickness Standards
Brinell Scopes ¹	(1 to 6) mm	11 μm	Stage Micrometer

Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Bench and Floor Scales ¹	(0.001 to 5 000) lb	0.000 7 lb/lb	NIST 105 Class F Weights NIST Handbook 44
Analytical Balances ¹	(0.001 mg to 13 kg)	0.19 μg/g	ASTM E617 Class 1 Weights NIST Handbook 44
Pressure ¹	(-13 to 300) psi (300 to 1 000) psi	0.1 psi 1.3 psi	Pressure Calibrator
	(1 000 to 10 000) psi	3.9 psi	Pressure Transducers
Environmental Pressure Gage ¹	(0 to 100) inH ₂ O	0.017 inH ₂ O	Pressure Module
Durometers ¹	Up to 100 duro	0.35 duro	Direct Verification per ASTM D2240 Durometer Calibrator
Spring Force	(0.1 to 45) N	0.05 N	Triple Beam Balance, Video Measuring Machine
Indenter Angle	(20 to 40) °	0.07 °	
Indenter Radius	(0.05 to 0.1) in	340 μin	
Indenter Length	(0.049 to 0.198) in	330 μin	Gage Blocks
Brinell Hardness Testers ¹ Verification of Test Force	(500, 750, 1 500, 2 000, 3 000) kgf	7.2 Kgf	Direct Verification per ASTM E10 using Morehouse Proving Ring
Verification of Indenter Mean Diameter	10 mm 5 mm	0.002 mm 0.002 mm	
Brinell Hardness Testers ¹	(1 to 7) mm	0.03 mm	Indirect Verification per ASTM E10 using Brinell Test Blocks & Brinell Scope



Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Knoop and Vickers Hardness Testers ¹	(1 to 200) μm (1 to 200) μm	0.25 μm 0.17 μm	Indirect Verification per ASTM E384 using Knoop & Vickers Test Blocks
Leeb Hardness Tester ¹	550 LD, 836 LD	20 LD	Indirect Verification per ASTM A596 using Leeb Test Block
Rockwell Hardness Testers ¹	HRA Low HRA Med HRA High HRBW Low HRBW Med HRBW High HRC Low HRC Med HRC High HRE Low HRE Med HRE High HRF Low HRF Med HRF High HRH Low HRH Med HRH High HRKW Low HRKW Med HRKW High HRMW Low HRMW Med HRMW High	1.2 HRA 1.2 HRA 0.75 HRA 1.4 HRBW 1.4 HRBW 1.3 HRBW 1.2 HRC 1.2 HRC 0.7 HRC 1.3 HRE 1.4 HRE 1.4 HRE 1.4 HRF 1.4 HRF 1.4 HRF 1.4 HRH 1.4 HRH 1.4 HRH 1.4 HRKW 1.3 HRKW 1.3 HRKW 1.4 HRMW 1.4 HRMW 1.3 HRMW	Indirect Verification per ASTM E18 using Rockwell Test Blocks



Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Rockwell Hardness Testers ¹	HR15N Low HR15N Med HR15N High	1.5 HR15N 1.3 HR15N 0.9 HR15N	Indirect Verification per ASTM E18 using Rockwell Test Blocks
	HR30N Low HR30N Med HR30N High	1.3 HR30N 1.3 HR30N 0.9 HR30N	
	HR45N Low HR45N Med HR45N High	1.4 HR45N 1.3 HR45N 0.95 HR45N	
	HR15TW Low HR15TW Med HR15TW High	2 HR15TW 1.4 HR15TW 1.5 HR15TW	
	HR30TW Low HR30TW Med HR30T High	2 HR30TW 1.5 HR30TW 1.3 HR30TW	
Rockwell Superficial Hardness Testers ¹	HR45TW Low HR45TW Med HR45TW High	2.0 HR45TW 1.3 HR45TW 1.4 HR45TW	Indirect Verification per ASTM E18 using Rockwell Test Blocks
Force ¹	(0.001 to 200) lb (200 to 10 000) lb (10 000 to 50 000) lb	0.05 % of reading 0.07 % of reading 0.1 % of reading	Dead Weight Load Cell Load Cell
Torque Tools ¹	4 lbf·in to 2 500 lbf·ft	0.3 % of reading	AKO Torque System
Viscosity Rotational Viscometers	500 cP 5 000 cP	0.02 cP/cP	Viscosity Solutions, Temperature Bath
Viscosity Cups	17.82 cP 65.28 cP 231 cP	0.03 cP/cP	Viscosity Solutions, Temperature Bath, Stopwatch ASTM D4212



Photometry and Radiometry

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Optical Power - Measure ¹ (800 to 1 650) nm	(-70 to +20) dBm	0.03 dB/dBm	Agilent 81533B Interface, 81525A Optical Head
Optical Power - Source ¹ (820, 1 310, 1 550) nm	(-60 to 0) dB	0.05 dB/dB	Agilent 81554SM Laser Source Module, 81533B Interface, 81525A Optical Head, 81655A Laser Module, 81570A Optical Attenuator, and 81578A Optical Attenuator
Optical Attenuation - Source ¹ (700 to 1 650) nm	(-60 to 0) dB	0.04 dB/dB	Agilent 81570A and 81578A Optical Attenuators
Optical Wavelength - Measure ¹	(700 to 1 650) nm	0.05 nm	Agilent 86120B Multi- Wavelength Meter

Thermodynamic

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Immersion Probes ¹	(-95 to 140) °C	0.03 °C	Fluke 9190A Drywell with PRT
Temperature - Measure ¹	(-30 to 600) °C	0.03 °C	Hart Scientific 1502 Indicator with PRT
System Accuracy Test ¹ (SAT)	(0 to 2 200) °F	2.6 °F	Certified Thermocouple
Temperature Uniformity Survey ¹ (TUS)	(0 to 2 200) °F	4.9 °F	MV 1000 Data Logger or Equivalent

Time and Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Time Interval ¹	(1 to 86 400) s	0.000 45 s	Agilent 53132A Counter & Spectracom 8197B GPS Oscillator

Time and Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Frequency Measure ¹	0.1 Hz to 225 MHz	6.7 parts in 10 ⁻¹¹ Hz	Agilent 53132A Counter & Spectracom 8197B GPS Oscillator
	0.1 Hz to 26.5 GHz	6.7 parts in 10 ⁻¹¹ Hz	Agilent N5531S Measuring Receiver, SRS FS725 Frequency Standard
Frequency Source ¹	10 MHz	6.7 parts in 10 ⁻¹¹ Hz	SRS FS725 Frequency Standard
	0.1 mHz to 40 GHz	6.7 parts in 10 ⁻¹¹ Hz	Agilent 3325B Function Generator, Agilent E8257D Signal Generator, SRS FS725 Frequency Standard
Tachometers ¹ Contact Non-Contact	(1 to 6 500) rpm (500 to 40 000) rpm	0.08 % of reading	King Nutronics 3711-B Tachometer Test Set
Tachometers ¹ Non-Contact	(0.01 to 100 000) rpm	0.005 % of reading	Fluke 5520A Multiproduct Calibrator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope
2. This scope is formatted as part of a single document including Certificate of Accreditation No. ACT-1272.03.



Vice President

